



National Voluntary Laboratory Accreditation Program



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

Oak Ridge Metrology Center
P. O. Box 2009
Oak Ridge, TN 37831-8091
W. T. (Bill) McKeethan
Phone: 865-574-2707 Fax: 865-574-2802
E-mail: mckeethanwt@y12.doe.gov

CALIBRATION LABORATORIES

NVLAP LAB CODE 105000-0

Scope Revised: 2006-12-18

DIMENSIONAL

NVLAP Code: 20/D03

Gage Blocks, Steel and Chrome Only

Range	Best Uncertainty (\pm) ^{note 1}	Remarks
0.010 in to 0.090 in	2.7 μ in	Mechanical Comparison
0.01 in to 1.000 in	2.1 μ in	Mechanical Comparison
2.0 in to 4.0 in	(3.1 + 2.0L) μ in ^{note 3}	Mechanical Comparison
0.20 mm to 2.5 mm	0.069 μ m	Mechanical Comparison
2.60 mm to 25 mm	0.055 μ m	Mechanical Comparison
30 mm to 100 mm	(0.08 + 2.0L) μ m ^{note 4}	Mechanical Comparison

NVLAP Code: 20/D05

Length

Range in m	Best Uncertainty (\pm) ^{notes 1, 4}	Remarks
0 to 1.2	(0.3 + 0.4L) μ m	Step and End Gages using M-48 Coordinate Measuring Machine

2006-04-01 through 2007-03-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 105000-0

Scope Revised: 2006-12-18

NVLAP Code: 20/D06

Line Standards

Range in mm

0 to 800

Best Uncertainty (\pm) ^{notes 1, 4}

(0.36 + 0.68L) μm

Remarks

CMM (optical)

NVLAP Code: 20/D08

Optical Grid Plates/Reference Planes

Range

Field of view ^{note 5}

Best Uncertainty (\pm) ^{notes 1, 4}

0.2 μm

Remarks

CMM (optical), Measurements
taken within camera field of view

Range in mm

0 to 848

Best Uncertainty (\pm) ^{notes 1, 4}

(0.40 + 0.61L) μm

Remarks

CMM (optical), maximum length
and width (600 x 600) mm

NVLAP Code: 20/D09

Roundness

Range

to 6 inches Diameter and 4 inches
Height

Best Uncertainty (\pm) ^{note 1}

0.1 μm

Remarks

Roundness Instrument

NVLAP Code: 20/D12

Surface Texture

Range

41 μin to 120 μin (1.04 μm to 3.05 μm)

13 μin to 40 μin (0.33 μm to 1.02 μm)

12 μin (0.31 μm)

Best Uncertainty (\pm) ^{note 1}

5.03 μin (0.13 μm)

1.74 μin (0.044 μm)

0.85 μin (0.021 μm)

Remarks

Ra (Roughness Average)

Ra (Roughness Average)

Ra (Roughness Average)

NVLAP Code: 20/D15

Two Dimensional Gages

Range in m

0.8 x 1.2

Best Uncertainty (\pm) ^{notes 1, 4}

(0.45 + 0.6L) μm

Remarks

M-48 CMM

2006-04-01 through 2007-03-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 105000-0

Scope Revised: 2006-12-18

NVLAP Code: 20/D18

Gears

Range	Best Uncertainty (\pm) ^{note 1}	Remarks
to 6 inches Diameter	0.9 μm	Involute Profile
to 6 inches Diameter and Infinite Lead	0.8 μm	Helix
to 6 inches Diameter and 99 inches Lead	0.9 μm	Helix
to 6 inches Diameter and 32 inches Lead	1.1 μm	Helix
to 6 inches Diameter and 16 inches Lead	1.2 μm	Helix
to 6 inches Diameter and 11 inches Lead	1.4 μm	Helix
to 6 inches Diameter (pin offset)	0.7 μm	Pin Master
to 6 inches Diameter (pin diameter)	0.5 μm	Pin Master
to 6 inches Diameter (pin roundness)	0.3 μm	Pin Master
to 24 inches Diameter	1.6 arcseconds	Index and Runout

TIME AND FREQUENCY

NVLAP Code: 20/F01

Frequency Dissemination

Range	Best Uncertainty (\pm) ^{note 1}	Remarks
1 MHz, 5 MHz, 10 MHz	8.75×10^{-12}	NIST FMAS

THERMODYNAMIC

NVLAP Code: 20/T05

Pressure

Range	Best Uncertainty (\pm) ppm ^{note 1}	Remarks
Pneumatic Deadweight Piston Gauge (absolute Mode) - Direct Pressure Comparison		
1.2 psia to 23.6 psia (8.3 kPa to 162.7 kPa)	500	
5.7 psia to 95.6 psia (39.3 kPa to 659.1 kPa)	101	
41.9 psia to 1001.6 psia (288.9 kPa to 6905.8 kPa)	45	

2006-04-01 through 2007-03-31

Effective dates

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 105000-0

Scope Revised: 2006-12-18

Pneumatic Deadweight Piston Gauge (Gauge Mode) - Direct Pressure Comparison

1.2 psia to 23.6 psia (8.3 kPa to 162.7 kPa)	26	Nitrogen
5.7 psia to 95.6 psia (39.3 kPa to 659.1 kPa)	22	Nitrogen
41.9 psia to 1001.6 psia (288.9 kPa to 6905.8 kPa)	43	Nitrogen

Hydraulic Deadweight Piston Gauge (Gauge Mode) - Direct Comparison

203 psig to 3771 psig (1.4 MPa to 26 MPa)	60	Oil
2031 psig to 19 870 psig (14 MPa to 137 MPa)	70	Oil
4061 psig to 39 595 psig (28 MPa to 273 MPa)	70	Oil

NVLAP Code: 20/T07

Resistance Temperature Devices

Range in °C	Best Uncertainty (\pm)^{note 1}	Remarks
0.01 to 29.7646	0.0017 °C	Comparison

1. Represents an expanded uncertainty using a coverage factor, $k = 2$, at an approximate level of confidence of 95 %.
2. Realizable uncertainty depends on frequency being measured, customer requirements, and suitability of customer's equipment.
3. L is length in inches.
4. L is length in meters.
5. Glass Reticles, Stage Micrometers, Glass Magnification Scales, Orthogonality Standards, and Calibration Charts.

2006-04-01 through 2007-03-31

Effective dates

For the National Institute of Standards and Technology